



Shenzhen GCT Technical Service Co.,Ltd.

HOTLINE: 400-096-8018

WWW.GCT-LAB.COM

Certificate of Compliance

Certificate No. : GCT200378083E
Applicant : Shenzhen YKD Technology Co.,Ltd.
Applicant Address : 6F,Building 6. Huaxinruiming Industrial Area,Langrong Rd,Xinwei Village,Dalang Community,Dalang Street, Longhua,Shenzhen,China
Manufacturer : Shenzhen YKD Technology Co.,Ltd.
Manufacturer Address : 6F,Building 6. Huaxinruiming Industrial Area,Langrong Rd,Xinwei Village,Dalang Community,Dalang Street, Longhua,Shenzhen,China
Product : Finger Oximeter
M/N : X004C
X004B,X002
Trademark : N/A
Test Standard : EN 60601-1-2:2015
(CISPR 11:2015;IEC 61000-3-2:2014;IEC 61000-3-3:2013;
IEC 61000-4-2:2008;IEC 61000-4-3:2006;IEC 61000-4-4:2012;
IEC 61000-4-5:2014;IEC 61000-4-6:2013;IEC 61000-4-8:2009;
IEC 61000-4-11:2004)

The EUT described above has been tested by us with the listed standards and found in compliance with the council MDD directive 93/42/EEC. It is possible to use CE marking to demonstrate the compliance with this MDD Directive. It is only valid in connection with the test report number: GCT200378083ER.

CE



Manager

Issue Date: Apr. 03, 2020